OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-19.49
Project number 2305042

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Changzhou Sunmi Scale Technology Co.,ltd. Xinzha Technology Industrial Park Buidling 19&21, 213000 Changzhou, Jiangsu China
Identification of the certified type	A **Non-automatic weighing instrument** Type : S2
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III) or (IIII)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
27 September 2019

C. Oosterman
Head Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

OIML Certification System

INSPECTION RvA I 122

OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-19.49
Project number 2305042

The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Report:

- No. NMi-2305042-01 dated 27 September 2019 that includes 63 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class		(III) or (IIII)
Maximum capacity		6 kg ≤ Max ≤ 30 kg
Verification scale interval		e ≥ 2 g
Weighing range(s)		Single interval Multi-interval
Maximum number of scale intervals		n ≤ 3000
Maximum number of weighing ranges		2
Tare		T ≤ -50% + e for instruments with one weighing range T ≤ -Max_1 + e_1 for multi-interval instruments
Temperature range		-10 °C / +40 °C
Power supply voltage		100 – 240 V AC 50/60 Hz
Application		Intended to be used for direct sales to the public
Software identification	Version number	10xxx (x = 0… 9)